

**Calibration Report n°**

**ABRXXXXXX\_55586**

**Issued**

**10/02/2026**

### Customer

Name CUSTOMER  
Address ADDRESS  
ADDRESS  
Country COUNTRY

### Order

Number 4500134423

### Instrument

Type ABRASION CHECK  
Model ABRASION CHECK  
Producer GIBITRE INSTRUMENTS SRL  
Serial Number ABRXXXXXX

### Calibration

Date of the measures **12/01/2026**  
Technician **Alan Arsuffi**

[Habilitation for Calibration](#)

### Reference Standard

The calibration is made in accordance to the requirements of the following standards:

**ISO 4649: Rubber, vulcanized or thermoplastic— Determination of abrasion resistance using a rotating cylindrical drum device**

The measurement uncertainties stated in this document have been determined according to the ISO/IEC Guide 98 and to EA-4/02. Usually they have been estimated as expanded uncertainty obtained multiplying the standard uncertainty by the coverage factor k corresponding to a confidence level of about 95%. Normally, this factor k is 2.


Calibration made by:

Alan Arsuffi



Calibration Report issued by:

Ivan Locatelli



**Calibration Report n°**
**ABRXXXXX\_55586**

The measurement results reported in this Calibration Report were obtained following the procedures given in the following pages, where the reference standards or instruments are indicated which guarantee the traceability chain of the laboratory, and the related calibration certificates in the course of validity are indicated as well. They relate only to the calibrated item and they are valid for the time and conditions of calibration, unless otherwise specified.

Reference Instruments	Producer	Serial N.	Gibitre Code	Certificate N.	Calibration Laboratory	Issue Date	Due Date	Uncertainty	Unit
Digital Caliper	Mitutoyo	1019218	CLB02	<a href="#">LAT 051 CT-CC-0227-2022</a>	TRESCAL	23/05/22	23/05/2027	0,01	mm
Chronometer	RS COMPONENTS	GBT.CN.01/13	CRO02 [0-60 s]	<a href="#">LAT 056 23-0199 2023</a>	GAMMA MISURE	09/02/23	09/02/2028	0,10	s
Standard rubber Abrasimeter	Cerisie	14114-14214	GNA01	<a href="#">RP 823-2024</a>	CERISIE	21/10/24	21/10/2026	0,00	g
Set of weights 1g - 5 Kg	Sartorius AG	PES01	PES01 [1-500 g]	<a href="#">LAT 117 23 M 143 I</a>	CIBE	20/04/23	20/04/2026	0,00	g
Set of weights 1g - 5 Kg	Sartorius AG	PES01	PES01 [1000-5000 g]	<a href="#">LAT 117 23 M 143 I</a>	CIBE	20/04/23	20/04/2026	0,00	g
Chronometer + Calibrator	Gibitre Instruments srl	C5-CH-1	C5-GB3-CAL-1 + C5-CH-1	<a href="#">CAL105 32048</a>	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,16	s
Digital Calliper	Mitutoyo	A16222743	C5-CL-200	<a href="#">CAL105 32048</a>	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,01	mm
Load Cell for Abrasion Tester + Calibrator	DS Europe	166166	C5-GB3-CAL-1 + C5-AB-1	<a href="#">CAL105 32048</a>	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,03	N
Calibrator + 22 N Load Cell	Interface + Gibitre Instruments	C5-HI-1	C5-GB3-CAL-1 + C5-HI-1	<a href="#">CAL105 32048</a>	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,00	N

**ENVIRONMENTAL CONDITIONS**

Room Temperature	(23 ± 2) °C
Relative Humidity	(50 ± 10) %



## Calibration Report n°

## ABRXXXXXX\_55586

Calibration of: **Vertical force applied to the sample**

Procedure: The vertical forces applied using the weights supplied are measured using the reference force sensor

Reference Standard: **ISO 4649 Par. 5.1**

Reference Instruments:

**C5-GB3-CAL-1 + C5-AB-1**                      Uncertainty: 0,0300    N                      Deviation 0,01            N

Expected Force	Minimum Allowed	Maximum Allowed	Measure 1	Measure 2	Measure 3	Mean	Accuracy	Uncertainty U_ext_95%	Outcome
N	N	N	N	N	N	N	N	N	
5	4,9	5,1	5,1	5,1	5,1	5,1	0,1	0,030	ok
10	9,8	10,2	10,2	10,2	10,2	10,2	0,2	0,030	ok

Calibration of: **Rotation Frequency**

Procedure: The rotation frequency is calibrated by measuring, with the reference Chronometer, the time to make the complete test cycle

Reference Standard: **ISO 4649 Par. 5.1**

Reference Instruments:

**C5-GB3-CAL-1 + C5-CH-1**                      Uncertainty: 0,1618    s                      Deviation 0,00            s

Rotation Frequency	Time for 84 Revolutions	Minimum Allowed	Maximum Allowed	Measure 1	Measure 2	Measure 3	Mean	Accuracy	Uncertainty U_ext_95%	Outcome
Rev/min	Sec	Sec	Sec	Sec	Sec	Sec	Sec	Sec	Sec	
40	126	122,9	129,2	127,9	127,1	127,8	127,6	1,6	0,524	ok

**Calibration Report n°**

**ABRXXXXX\_5586**

Calibration of: **Number of Rotations**

Procedure: The number of rotations before automatic stop is counted

Reference Standard: **ISO 4649 Par. 5.1**

Number of Rotations num	Minimum Allowed num	Maximum Allowed num	Measure 1 num	Measure 2 num	Measure 3 num	Outcome
42	41	43	42,0	42,0	42,0	ok
84	83	85	84,0	84,0	84,0	ok

Calibration of: **Lateral Displacement**

Procedure: The lateral displacement of the sample holder is measured using the reference Calliper after a complete cycle of 42 rotations

Reference Standard: **ISO 4649 Par. 5.1**

Reference Instruments:

**C5-CL-200**                      Uncertainty: 0,0148 mm      Deviation 0,01 mm

Expected displacem. mm/Rev	Displacem. for 42 Rotations mm	Minimum Allowed mm	Maximum Allowed mm	Measure 1 mm	Measure 2 mm	Measure 3 mm	Mean mm	Accuracy mm	Uncertainty U_ext_95% mm	Outcome
4,2	176,4	175,6	177,2	176,70	176,60	176,60	176,63	0,23	0,07	ok

**Calibration Report n°**

**ABRXXXXXX\_55586**

Calibration of: **Abrasiveness of Abrasive Sheet**

Procedure: Perform 3 complete tests without sample rotation using Reference samples. Weigh the samples before and after the tests with the reference scale

Reference Standard: **ISO 4649 Par. 5.2**

Reference Instruments:

**GNA01**

**C5-GB3-CAL-1 + C5-HI-1**

Uncertainty: 0,0020 N

Deviation 0,01 N

Expected Mass Loss mg	Minimum Allowed mg	Maximum Allowed mg	Measure 1 mg	Measure 2 mg	Measure 3 mg	Mean mg	Uncertainty U_ext_95% mg	Outcome
200	180	220	203,0	204,0	201,0	202,7	1,764	ok

**Δm<sub>r</sub>:** Reference Mass Loss of the refer. compound = **202,7**



**Gibitre Instruments S.r.l.**

Via Dell'Industria,73 24126 BERGAMO - ITALY

Tel. +39035460146

E-mail: info@gibitre.it - <http://www.gibitre.it>

CER\_ABR\_001

Rev: 01

IN

**Calibration Report n°**

**ABRXXXXXX\_55586**

**Verification of instrument safety devices**

✓	Verification that the instrument's safety devices are working properly
---	--

✓	Final verification that the instrument is functioning properly with safety devices active
---	---