

CER_CSET_001	Rev: 01	IN
--------------	---------	----

<b>Calibration Report n°</b>	<b>C-SETXXXXXX_66060</b>
<b>Issued</b>	<b>08/05/2026</b>

**Customer**

Name	CUSTOMER
Address	ADDRESS
	ADDRESS
Country	COUNTRY

**Order**

Number

**Instrument**

Type	COMPRESSION SET
Model	COMPRESSION SET - PV3307
Producer	GIBITRE INSTRUMENTS S.R.L.
Serial Number	C-SETXXXXXX

**Calibration**

Date of the measures	<b>08/05/2026</b>	
Technician	<b>Mirko Persico</b>	<a href="#">Habilitation for Calibration</a>

**Reference Standard**

The calibration is made in accordance to the requirements of the following standards:  
**ISO 815: Rubber, vulcanized or thermoplastic — Determination of compression set — Part 1: At ambient or elevated temperatures**

**ISO 18899: Rubber - Guide to the calibration of test equipment**

The measurement uncertainties stated in this document have been determined according to the ISO/IEC Guide 98 and to EA-4/02. Usually they have been estimated as expanded uncertainty obtained multiplying the standard uncertainty by the coverage factor k corresponding to a confidence level of about 95%. Normally, this factor k is 2.


Calibration made by:

**Mirko Persico**



Calibration Report approved by:

**Ivan Locatelli**



**Calibration Report n°**
**C-SETXXXXXX\_66060**

The measurement results reported in this Calibration Report were obtained following the procedures given in the following pages, where the reference standards or instruments are indicated which guarantee the traceability chain of the laboratory, and the related calibration certificates in the course of validity are indicated as well. They relate only to the calibrated item and they are valid for the time and conditions of calibration, unless otherwise specified.


Reference Instruments	Producer	Serial N.	Gibitre Code	Certificate N.	Calibration Laboratory	Issue Date	Due Date	Uncertainty	Unit
Calibration Plate PS 10	Mahr	11654	7053546	<a href="#">LAT 052 2504872DSI</a>	KIWA CERMET ITALIA	26/03/25	26/03/2030	0,7600	µm
Roughness Tester	Mahr	01770/19	RUG_01	<a href="#">RUG_01_23</a>	GIBITRE INSTRUMENTS	11/04/23	11/04/2028	0,7800	µm

## ENVIRONMENTAL CONDITIONS

Room Temperature	(23 ± 2) °C
Relative Humidity	(50 ± 10) %

**Calibration Report n°**

**C-SETXXXXXX\_66060**



Calibration of: **Surface Rugosity**  
 Standards: **PV 3307 Elastomer Components - Plastic and Elastic Deformability**

Procedure: The calibration is performed by testing the rugosity in different points of the surfaces of the instrument which are in contact with the sample. The measure is performed using a calibrated Rugosimeter

Reference Instruments:

RUG\_01                      Uncertainty: (Rz)                      0,7800                      μm                      Deviation                      0,00                      μm

Surface	Max permitted value Rz [μm]	Measure 1 Rz [μm]	Measure 2 Rz [μm]	Measure 3 Rz [μm]	Max Rz [μm]	Uncertainty U_ext_95% μm	Outcome μm
Z	6,3000	3,420	3,774	3,554	3,7740	0,8068	ok
W	6,3000	3,702	3,668	3,818	3,8180	0,7853	ok
1	6,3000	4,597	4,696	4,807	4,8070	0,7894	ok
2	6,3000	4,906	5,027	5,553	5,5530	0,8753	ok
3	6,3000	5,296	5,208	5,136	5,2960	0,7855	ok
4	6,3000	4,807	5,323	5,249	5,3230	0,8439	ok
5	6,3000	4,373	4,627	3,697	4,6270	0,9573	ok
6	6,3000	5,625	5,667	4,445	5,6670	1,1181	ok