

Calibration Report n°

FLAXXXXXX_65343

Issued

12/03/2026

Customer

Name CUSTOMER
Address ADDRESS
ADDRESS
Country COUNTRY

Order

Number 26/01267

Instrument

Type FLAME TESTER
Model FLAMMABILITY CHECK
Producer GIBITRE INSTRUMENTS SRL
Serial Number FLAXXXXXX

Calibration

Date of the measures **26/02/2026**
Technician **Alan Arsuffi** [Habilitation for Calibration](#)

Reference Standard

The calibration is made in accordance to the requirements of the following standards:

ASTM D 5207: Confirmation of 20 and 125 mm Test Flames for Small-Scale Burning Tests on Plastic Materials


UL 94: Tests for flammability of plastic materials for parts in devices and appliances

ASTM D 5025: Laboratory Burner Used for Small-Scale Burning Tests on Plastic Materials

The measurement uncertainties stated in this document have been determined according to the ISO/IEC Guide 98 and to EA-4/02. Usually they have been estimated as expanded uncertainty obtained multiplying the standard uncertainty by the coverage factor k corresponding to a confidence level of about 95%. Normally, this factor k is 2.

Calibration made by:

Calibration Report approved by:

Ivan Locatelli


The measurement results reported in this Calibration Report were obtained following the procedures given in the following pages, where the reference standards or instruments are indicated which guarantee the traceability chain of the laboratory, and the related calibration certificates in the course of validity are indicated as well. They relate only to the calibrated item and they are valid for the time and conditions of calibration, unless otherwise specified.

Calibration Report n°
FLAXXXXXX_65343

| Reference Instruments | Producer | Serial N. | Gibitre Code | Certificate N. | Calibration Laboratory | Issue Date | Due Date | Uncertainty | Unit |
|---------------------------------------|-------------------------|----------------|-------------------------|---|------------------------|------------|------------|-------------|------|
| Digital thermometer with Pt100 sensor | Delta Ohm | 12036501-13023 | TER-12036501-13023 | LAT 051 C119206D40 | TRESCAL | 15/05/19 | 15/05/2026 | 1,90 | °C |
| Chronometer | RS COMPONENTS | GBT.CN.01/13 | CRO02 [0-60 s] | LAT 056 23-0199 2023 | GAMMA MISURE | 09/02/23 | 09/02/2028 | 0,10 | s |
| Digital Caliper | Mitutoyo | 1019218 | CLB02 | LAT 051 CT-CC-0227-2022 | TRESCAL | 23/05/22 | 23/05/2027 | 0,01 | mm |
| Optical Coordinate Measuring Machine | OGP HOMMEL Italia srl | SKL2252499 | MDI02 [Ø 0,5-5 mm] | MDI02 21172 | GIBITRE INSTRUMENTS | 26/02/25 | 26/02/2028 | 0,00 | mm |
| Chronometer + Calibrator | Gibitre Instruments srl | C5-CH-1 | C5-GB3-CAL-1 + C5-CH-1 | CAL105 32048 | GIBITRE INSTRUMENTS | 02/09/25 | 02/09/2026 | 0,16 | s |
| Thermocouple + Calibrator | Gibitre Instruments srl | C5-T-TC1 | C5-GB3-CAL-1 + C5-T-TC1 | CAL105 32048 | GIBITRE INSTRUMENTS | 02/09/25 | 02/09/2026 | 3,66 | °C |
| Digital Calliper | Mitutoyo | A16222743 | C5-CL-200 | CAL105 32048 | GIBITRE INSTRUMENTS | 02/09/25 | 02/09/2026 | 0,01 | mm |

ENVIRONMENTAL CONDITIONS

| | |
|-------------------|-------------|
| Room Temperature | (23 ± 2) °C |
| Relative Humidity | (50 ± 10) % |

| | | |
|-------------|---------|----|
| CER_FLA_001 | Rev: 02 | IN |
|-------------|---------|----|

Calibration Report n°

FLAXXXXXX_65343

Calibration of: **20-mm Test Flame - time for the temperature to rise from 100 to 700°C**

Procedure: The calibration is performed using a thermocouple with a 1.75g copper tip. The burner must be set with flame with 20 mm height and blue cone. The flame must stabilize 5 minutes before the start of the calibration test. The thermal probe must be placed over the flame at 10 mm distance from the top of the burner. The timer must be started when the temperature is (100 +/- 2) °C and must be stopped when the temperature is (700 +/- 3)°C. The time required to reach the temperature must be (44 +/-2) seconds

Reference Standard: **ASTM D 5207**

Reference Instruments:

C5-GB3-CAL-1 + C5-T-TC1 Uncertainty: 3,6575 *C Deviation 1,00 *C
C5-GB3-CAL-1 + C5-CH-1 Uncertainty: 0,1618 s Deviation 0,00 s

| Set Value | Minimum Allowed | Maximum Allowed | Measure 1 | Measure 2 | Measure 3 | Mean | Accuracy | Uncertainty U_ext_95% | Outcome |
|-----------|-----------------|-----------------|-----------|-----------|-----------|-------|----------|-----------------------|---------|
| s | s | s | s | s | s | s | s | s | |
| 44 | 42 | 46 | 45,7 | 45,4 | 45,9 | 45,67 | 1,67 | 0,333 | ok |

Calibration Report n°

FLAXXXXXX_65343

Calibration of: **Time measuring device of the system**

Sensor Type: **Timer**

Resolution: **1 s**

Procedure: After the start of the test, the time reading of the instrument is compared after 60 seconds with the reading of the reference chronometer (see page 2).

Reference Standard: **UL 94**

Reference Instruments:

C5-GB3-CAL-1 + C5-CH-1 Uncertainty: **0,1618 s** Deviation **0,00 s**

| Instrument Reading | Minimum allowed | Maximum Allowed | Reference Reading 1 | Reference Reading 2 | Reference Reading 3 | Mean | Error | Uncertainty U_ext_95% | Outcome |
|--------------------|-----------------|-----------------|---------------------|---------------------|---------------------|-------|-------|-----------------------|---------|
| s | s | s | s | s | s | s | s | s | |
| 60 | 59 | 61 | 60,11 | 60,24 | 59,96 | 60,10 | 0,10 | 0,62 | ok |

Calibration of: **Templates for sample positioning**

Procedure: The templates are measured using the reference calliper

Reference Standard: **UL 94**

Reference Instruments:

C5-CL-200 Uncertainty: **0,0148 mm** Deviation **0,01 mm**

| Expected Size | Minimum Allowed | Maximum Allowed | Measure 1 | Measure 2 | Measure 3 | Mean | Accuracy | Uncertainty U_ext_95% | Outcome |
|---------------|-----------------|-----------------|-----------|-----------|-----------|-------|----------|-----------------------|---------|
| mm | mm | mm | mm | mm | mm | mm | mm | mm | |
| 10 | 9 | 11 | 10,54 | 10,56 | 10,56 | 10,55 | 0,55 | 0,020 | ok |
| 12 | 11 | 13 | 12,96 | 12,98 | 12,99 | 12,98 | 0,98 | 0,023 | ok |

Calibration Report n°

FLAXXXXXX_65343

Verification of instrument safety devices

| | |
|---|---|
| ✓ | Functional verification of Bunsen safety thermocopy |
|---|---|

| | |
|---|---|
| ✓ | Check that gas supply pipe is not expired |
|---|---|

| | |
|---|---|
| ✓ | Final verification that the instrument is functioning properly with safety devices active |
|---|---|

Disclaimer

The outcome of verifications regarding the proper functioning of safety devices on the instrument refers to the time when the verification is performed.

Periodic verification of the maintenance of the safety features of the instrument between successive maintenance services is the responsibility of the customer.