

Calibration Report n°**LTCXXXXXX_65432****Issued****23/04/2026****Customer**

Name CUSTOMER
Address ADDRESS
Address ADDRESS
Country COUNTRY

Order

Number RMG043136

Instrument

Type LOW TEMPERATURE CHECK
Model LOW TEMPERATURE CHECK (MACCHINA BASE)
Producer GIBITRE INSTRUMENTS SRL
Serial Number LTCXXXXXX

Calibration

Date of the measures **09/04/2026**
Technician **Francesco Ajazi** [Habilitation for Calibration](#)

Reference Standard

The calibration is made in accordance to the requirements of the following standards:


ISO 2921: Rubber, vulcanized -- Determination of low-temperature retraction (TR test)

ISO 812: Rubber, vulcanized or thermoplastic -- Determination of low-temperature brittleness

The measurement uncertainties stated in this document have been determined according to the ISO/IEC Guide 98 and to EA-4/02. Usually they have been estimated as expanded uncertainty obtained multiplying the standard uncertainty by the coverage factor k corresponding to a confidence level of about 95%. Normally, this factor k is 2.


Calibration made by:

Francesco Ajazi



Calibration Report approved by:

Ivan Locatelli



Calibration Report n°
LTCXXXXX_65432

The measurement results reported in this Calibration Report were obtained following the procedures given in the following pages, where the reference standards or instruments are indicated which guarantee the traceability chain of the laboratory, and the related calibration certificates in the course of validity are indicated as well. They relate only to the calibrated item and they are valid for the time and conditions of calibration, unless otherwise specified.

Reference Instruments	Producer	Serial N.	Gibitre Code	Certificate N.	Calibration Laboratory	Issue Date	Due Date	Uncertainty	Unit
PT100 Thermoresistance + Calibrator	Gibitre Instruments srl	C1-T-PTA	C1-GB3-CAL-1 + C1-T-PTA(T<0°C)	LAT 128T 141 7 22	ELLAB S.r.l.	25/09/22	25/09/2027	0,05	°C
Oscilloscope	Tektronic	B022027	OSC01 [5-2000 ms]	LAT 101 0719 2023 AC CR EO	TESI S.R.L.	04/10/23	04/10/2028	0,01	ms
Chronometer	RS COMPONENTS	GBT.CN.01/13	CRO02 [60-54000 s]	LAT 056 23-0199 2023	GAMMA MISURE	09/02/23	09/02/2028	0,15	s
Digital Caliper	Mitutoyo	1019218	CLB02	LAT 051 CT-CC-0227-2022	TRESCAL	23/05/22	23/05/2027	0,01	mm
Optical Coordinate Measuring Machine	OGP HOMMEL Italia srl	SKL2252499	MDI02 [Ø 0,5-5 mm]	MDI02 21172	GIBITRE INSTRUMENTS	26/02/25	26/02/2028	0,00	mm
Gauge Block 0.14mm.	Mitutoyo	950016	BPP02	LAT 051 C1242D4FA0	TRESCAL	31/01/24	31/01/2028	0,00	mm
Gauge Block 0.31mm.	Mitutoyo	950504	BPP03	LAT 051 C1242D4FA0	TRESCAL	31/01/24	31/01/2028	0,00	mm
Gauge Block 2mm.	MG	9765	BPP04	LAT 051 C1242D4F90	TRESCAL	31/01/24	31/01/2028	0,00	mm
Gauge Block 25 mm	MG	9510	BPP08	LAT 051 C1242D4F90	TRESCAL	31/01/24	31/01/2028	0,00	mm
Gauge Block 0,5 mm	Mitutoyo	130191	BPP42	LAT 051 C1232CA5F0	TRESCAL	19/10/23	19/10/2028	0,00	mm
Cylinder Gauge 0,400 mm	MG	04/633	SPI01	LAT 136 0178 CL 23	BOCCHI	28/03/23	28/03/2026	0,00	mm
Cylinder Gauge 1,25 mm	MG	04/634	SPI02	LAT 136 0179 CL 23	BOCCHI	28/03/23	28/03/2026	0,00	mm
Cylinder Gauge 2,500 mm	MG	04/635	SPI03	LAT 136 0180 CL 23	BOCCHI	28/03/23	28/03/2026	0,00	mm
Cylinder Gauge 25 mm	Mitutoyo	AA501	TAM01	LAT 136 0181 CL 23	BOCCHI	28/03/23	28/03/2026	0,00	mm
Electronic balance CP8021-OCE	Sartorius	13713388	BIL02 [0-8200 g]	LAT 134 B 022-2025	SOCIETA BILANCIAl PORRO	26/01/25	26/01/2027	0,08	g

Calibration Report n°
LTCXXXXX_65432

Chronometer + Calibrator	Gibitre Instruments srl	C13-CH-1	C13-GB3-CAL-1 + C13-CH-1	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,16	s
Weight	Gibitre Instruments srl	C13-WT-LT1	C13-WT-LT1	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,00	g
Weight	Gibitre Instruments srl	C13-WT-LT2	C13-WT-LT2	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,00	g
Thickness Block	Gibitre Instruments srl	C13-L-50	C13-L-50	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,01	mm
Thickness Block	Gibitre Instruments srl	C13-L-100	C13-L-100	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,01	mm
Thickness Block	Gibitre Instruments srl	C13-L-150	C13-L-150	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,01	mm
Digital Calliper	Mitutoyo	B18012619	C13-CL-200	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,01	mm
Thickness Block	Gibitre Instruments srl	C13-L-6.1	C13-L-6.1	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,01	mm
Thickness Block	Gibitre Instruments srl	C13-L-6.7	C13-L-6.7	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,01	mm
PT100 Thermoresistance + Calibrator	Gibitre Instruments srl	C13-T-PTA	C13-GB3-CAL-1 + C13-T-PTA(T<0°C)	CAL113_19653	GIBITRE INSTRUMENTS	24/09/24	24/09/2025	0,05	°C
Quick-entry time readers	Gibitre Instruments srl	C13-QI-1	C13-GB3-CAL-1 + C13-QI-1	CAL113_32198	GIBITRE INSTRUMENTS	02/09/25	02/09/2026	0,00	s

ENVIRONMENTAL CONDITIONS

Room Temperature	(23 ± 2) °C
Relative Humidity	(50 ± 10) %

Calibration Report n°
LTCXXXXXX_65432

Calibration of: **Temperature Measuring Device**
 Sensor Type: **Thermoresistance**
 Resolution: 0,1 °C

Procedure: The temperature of the bath is set at different values. The reading of the instrument is recorded in the first column. The readings of the reference thermometer (see page 2) are reported in the table below

Standard Reference: ISO 2921 - Annex A - Table A1 - Line 1
 ISO 812 - Annex B - Table B1- Line 8

Reference Instruments:
C13-GB3-CAL-1 + C13-T-PTA(T<0°C) Uncertainty: 0,0500 °C Deviation 0,50 °C

Instrument Reading °C	Minimum Allowed °C	Maximum Allowed °C	Reference Reading 1 °C	Reference Reading 2 °C	Reference Reading 3 °C	Mean °C	Error °C	Uncertainty U_ext_95% °C	Outcome
0	-0,5	0,5	-0,03	-0,05	-0,02	0,0	0,0	0,1	ok
-5	-5,5	-4,5	-5,07	-5,09	-5,05	-5,1	-0,1	0,1	ok
-35	-35,5	-34,5	-35,07	-35,05	-35,05	-35,1	-0,1	0,1	ok
-70	-70,5	-69,5	-70,05	-70,02	-70,02	-70,0	0,0	0,1	ok

Calibration of: **Heating of heat-transfer medium**
 Sensor Type: **Thermoresistance**
 Resolution: 0,1 °C

Procedure: After the start of the test, the temperature reading of the instrument is recorded with 10 minutes interval (time recoded using instrument clock). The temperature variation must be 10 ± 2 °C

Reference Standard: ISO 2921 - Annex A - Table A1 - Line 4

Reference Instruments:
C13-GB3-CAL-1 + C13-T-PTA(T<0°C) Uncertainty: 0,0500 °C Deviation 0,50 °C

Instrument Reading °C/10 min	Minimum Allowed °C/10 min	Maximum Allowed °C/10 min	Reference Reading 1 °C/10 min	Reference Reading 2 °C/10 min	Reference Reading 3 °C/10 min	Mean °C/10 min	Error °C/10 min	Uncertainty U_ext_95% °C/10 min	Outcome
10	8	12	10	10	10	10,0	0,0	0,1	ok

Calibration Report n°

LTCXXXXXX_65432

Calibration of: **Time mesuring device of the system**

Sensor Type: **PC Clock**

Resolution: **0,01 s**

Procedure: After the start of the test, the time reading of the instrument is compared after 60 seconds with the reading of the reference chronometer (see page 2).

Reference Standard: ISO 2921 - Annex A - Table A1 - Line 5

Reference Instruments:

C13-GB3-CAL-1 + C13-CH-1

Uncertainty: **0,1618 s**

Deviation **0,00 s**

Instrument Reading	Minimum allowed	Mazimum Allowed	Reference Reading 1	Reference Reading 2	Reference Reading 3	Mean	Error	Uncertainty U_ext_95%	Outcome
s	s	s	s	s	s	s	s	s	
60	59	61	60,00	60,00	60,00	60,00	0,00	0,16	ok

Calibration Report n°

LTCXXXXXX_65432

Calibration of: **Rack with test piece holders - Tension on the sample between 10 and 20 kPa**

Procedure: When the locking system is not active, each sample holder should move upwards when a force below 10 kPa is applied and should not move upwards when a force higher then 20 kPa is applied

Reference Standard: ISO 2921 - Annex A - Table A1 - Line 6

Reference Instruments:

C13-WT-LT1	Uncertainty: 0,0017 g	Deviation 0,01 g
C13-WT-LT2	Uncertainty: 0,0017 g	Deviation 0,00 g

Sample Holder	Force applied kPa	test 1	Test 2	Test 3
1	<10	ok	ok	ok
2	<10	ok	ok	ok
3	<10	ok	ok	ok
4	<10	ok	ok	ok
5	<10	ok	ok	ok
6	<10	ok	ok	ok

Sample Holder	Force applied kPa	test 1	Test 2	Test 3
1	>20	ok	ok	ok
2	>20	ok	ok	ok
3	>20	ok	ok	ok
4	>20	ok	ok	ok
5	>20	ok	ok	ok
6	>20	ok	ok	ok

Calibration Report n°
LTCXXXXX_65432

 Calibration of: **Retraction measuring devices - Accurate within $\pm 0,25$ mm**

 Sensor Type: **Encoder**
 Resolution: 0,16 mm

Procedure: Each sample holder is displaced by 100 mm using reference templates. The reading of the instrument is recorded in the table

 Reference Standard: **ISO 2921 - Annex A - Table A1 - Line 7**

Reference Instruments:

C13-L-50	Nom.Length:	50 mm	Uncertainty,0,0136	mm	Deviation	0,00	mm
C13-L-100	Nom.Length:	100 mm	Uncertainty,0,0136	mm	Deviation	0,00	mm
C13-L-150	Nom.Length:	150 mm	Uncertainty,0,0136	mm	Deviation	0,00	mm

Sample Holder	Set Displacem. mm	Minimum Allowed mm	Maximum Allowed mm	Instrument Reading 1 mm	Instrument Reading 2 mm	Instrument Reading 3 mm	Mean mm	Deviation mm	Uncertainty U_ext_95% mm	Outcome
1	50	49,75	50,25	50,04	50,13	50,04	50,07	0,07	0,11	ok
2	50	49,75	50,25	50,00	50,00	50,09	50,03	0,03	0,11	ok
3	50	49,75	50,25	50,04	49,96	49,96	49,99	-0,01	0,11	ok
4	50	49,75	50,25	49,91	50,00	50,00	49,97	-0,03	0,11	ok
5	50	49,75	50,25	50,03	50,05	50,10	50,06	0,06	0,10	ok
6	50	49,75	50,25	49,98	50,00	50,02	50,00	0,00	0,10	ok
1	100	99,75	100,25	100,00	100,00	100,09	100,03	0,03	0,11	ok
2	100	99,75	100,25	100,00	100,00	100,00	100,00	0,00	0,09	ok
3	100	99,75	100,25	100,00	100,00	100,00	100,00	0,00	0,09	ok
4	100	99,75	100,25	99,91	100,00	100,00	99,97	-0,03	0,11	ok
5	100	99,75	100,25	100,06	100,06	100,02	100,05	0,05	0,10	ok
6	100	99,75	100,25	100,00	100,02	100,04	100,02	0,02	0,10	ok

Calibration Report n°
LTCXXXXX_65432

 Calibration of: **Linear speed of striking edge for Brittleness point test**

Sensor Type: Limit Switches and internal timer

Resolution: 0,01 m/s

Procedure: Speed is measured by reading the time needed for the striker to move between the two measurement switches. The calibration is made by measuring the distance of the switches and the time needed for the displacement

Reference Standard: ISO 812 - Annex B - Table B1- Line 3

Reference Instruments:

C13-CL-200

Uncertainty: 0,0148 mm

Deviation 0,01 mm

Expected Distance of switches m	Instrument Reading 1 m	Instrument Reading 2 m	Instrument Reading 3 m	Mean m	Uncertainty U_ext_95% m
0,150	0,150	0,150	0,150	0,150	0,0000

Reference Instruments:

C13-GB3-CAL-1 + C13-QI-1

Uncertainty: 0,0006 s

Deviation 0,08 s

Expected Time Read. s	Instrument Reading 1 s	Instrument Reading 2 s	Instrument Reading 3 s	Mean s	Uncertainty U_ext_95% s
0,0753	0,0751	0,0751	0,0751	0,0751	0,0000

Instrument Speed Reading m/s	Minimum Allowed m/s	Maximum Allowed m/s	Measured Distance switches m	Measured Displacem. Time s	Measured Reference Speed m/s	Accuracy m/s	Uncertainty U_ext_95% m/s
2,0	1,8	2,2	0,150	0,0751	1,997	-0,003	0,0029

